



## Option 43 — Temperature Data

To access the Temperature Data menu in Display mode:

1. Press **D**.
2. Press **2** to invoke the Quick Jump option.
3. Press **43** for the **Temperature Data** menu, and then press **E**.

---

```
TEMPERATURE
87.21      DEGF
```

---

The Temperature prompt for the current process appears.

4. Press **P** to continue.

---

```
CORRECTION
FACTOR    1.000000
```

---

The Correction Factor prompt for the accumulated correction appears.

5. Press **H** to exit.

The main Display mode (DSP) prompt appears.